



भा.कृ.अनु.प. - केन्द्रीय कपास अनुसंधान संस्थान
पोस्ट बैग सं. 2, शंकरनगर पोस्ट ऑफिस, नागपुर -440 010
ICAR - CENTRAL INSTITUTE FOR COTTON RESEARCH
POST BAG NO 2, SHANKAR NAGAR POST OFFICE NAGPUR - 440 010(M.S) INDIA
Tel. Phone: 07103-275536, Fax: 07103-275529, EPBAX: 07103275637/38/39/49/17
Website: www.cicr.org.in, E-mail: cicrnagpur@gmail.com
{Location: Near Khapari Phata, Panjari, Wardha Road, Nagpur – 441 108}
ISO 9001-2008 Certified



F. No.(50)St/Equipment/Flame photometer/AM/HQ NGP/2017-18

Date: 24.01.2018

To,

Sub:- Quotation for the Purchase of **Flame Photometer with Compressor** – reg.

Sir,

Sealed quotations are invited by the Director, ICAR-Central Institute for Cotton Research, Nagpur for the purchase of **Flame Photometer with Compressor** as per the **Specification detailed in Annexure–A** on the terms and conditions prescribed below:

Terms & Conditions:

1. The rates quoted shall be for free delivery & installation at CICR including everything except GST may be shown separately. However, equipment quoted in foreign currency must be quoted on FOB basis, Custom Duty Exemption certificate will be issued only when the bid is quoted in foreign currency in case of foreign made item only & meant for CICR, Panjari, Near Khapari Phata, Wardha Road, Nagpur (Maharashtra).
2. The firm participating in the tender will be required to deposit an account payee bank draft of ₹ 5,000/- being the Earnest Money Deposit, in the name of "The Director, CICR, Nagpur" payable at Nagpur. Earnest Money in any other form shall not be accepted and the tender would be summarily rejected. The earnest money shall be refunded without interest after deposit of security money by the successful bidder. However, the successful bidder will be required to deposit security money calculated at **10% of the equipment cost** in the shape of Demand Draft in favour of "Director, CICR, Nagpur", payable at Nagpur which shall be refunded after six months of successful completion of the contract. **No interest shall be paid on EMD.**
3. Sealed quotation along with the literature and pamphlets should reach this office **on or before 20.02.2018 at 04.00 PM sharp and the quotations will be opened on 21.02.2018 at 02.00 PM.**
4. As the **Flame Photometer with Compressor** is urgently needed for our research work, the firms who will be able to make immediate supply **within 10 days of supply/purchase order** at Central Institute for Cotton Research, Panjari Farm, Wardha Road, Near Khapri, Nagpur should only quote the desired rates.

Contd..2...

-: 2 :-

5. Envelops may please be super-scribed with the word "**Flame Photometer with Compressor**".
6. Instrument indigenous or imported will be purchased from only those parties, who can ensure and undertake to provide the satisfactory after-sales-service within the country. The supplier should also ensure to offer a complete detailed package on back up support, prompt commissioning, after-sales-service including workshop backup, timely supply of **Flame Photometer with Compressor** to Scientists/Technicians on use.
7. Guarantee/Warranty period offered and after sales service available should be clearly mentioned.
8. The quotation should be quoted by solely local authorized stockiest Service distributor against this **Flame Photometer with Compressor**, who has necessary service backup facility.
9. Authorization Certificate should be enclosed with the quotation for this **Flame Photometer with Compressor**.
10. Those who are unable to provide the local service facility will not be entertained.
11. The make, quality, and specifications may be stated in detail/must be as per details given by this office (**catalogue may be enclosed**). In case supply is not made within a fixed period the office reserves the right to cancel the supply order.
12. The payment will be made only after completing delivery & installation according to the specifications, quality, and quantity given in the quotation/purchase order.
13. Other terms and conditions of supply if any may also be stated.
14. **Flame Photometer with Compressor** will be accepted only after giving satisfactory installation/ demonstrations.
15. The Director, ICAR-CICR, reserves the right to accept or reject one or all quotations without assigning any reason thereof.
16. The quotations, without the desired specification received after the due date is liable to be rejected.

Yours faithfully,

Encl: Annexure-A


21/1/18
Sr. Administrative Officer

Copy to:

1. Dr. A. Manikandan, Scientist, Crop Production Division, ICAR-CICR, Nagpur for information.

COMPANY LETTER HEAD**FORMAT OF QUOTATION**

Sl. No.	Technical Specification		Rate (₹)	Qty.	Amount (₹)
1.	Flame Photometer with Compressor			01 No.	
	Parameters	Specifications			
	Analysis	Calcium (Ca), Sodium (Na) & Potassium (K) and Lithium (Li) elements in soil and plant samples			
	Sensitivity	Na -0.5ppm, K-0ppm and Li -0.5ppm			
	Resolution	0.1 ppm			
	Power	220-240V			
	Fuel	Liquefied Petroleum Gas (LPG)			
	Interface	RS-232			
	Memory	Minimum storage of 99data			
	Additional accessories	1. Filters of Calcium (Ca), Lithium (Li) and Barium (Ba) 2. Professional PC software for Computer Interface 3. Portable data logger with Operating System – Windows, Display – 15 to 16 inch, RAM-3 to 4GB, Memory – 1TB			
				Total	

Authorized Signatory with seal/stamp

Address:

E-mail id:

Ph./Mob. No.